

# Column Care and Use Instructions

## YMC-Triart C18, C18 ExRS, C8, Phenyl, PFP, Bio C4

(HPLC: 5  $\mu\text{m}$ , 3  $\mu\text{m}$  / UHPLC: 1.9  $\mu\text{m}$ )

### 1. Introduction

Thank you for purchasing a YMC high-performance/ultra-high-performance liquid chromatography (HPLC/UHPLC) column. YMC-Triart is a multipurpose reversed phase column using hybrid silica gel. The column is designed to function under various analytical conditions and to deliver exceptional selectivity and stability.

YMC-Triart columns, which are manufactured under highly controlled conditions, must pass a series of strict tests before being accepted for shipment (Please refer to the column inspection report). To ensure optimal performance and durability of the column, please read these instructions carefully before using this column.


### 2. Specifications

Packing material	Particle size ( $\mu\text{m}$ )	Pore size (nm)	Functional group	End capped	Usable pH range	Usable temperature range	
						Regular use (recommended)	Upper limit
Triart C18	1.9, 3, 5	12	C18	Yes	1 – 12	20 – 40 °C	pH 1 – 7: 90 °C pH 7 – 12: 50 °C
Triart C18 ExRS		8					
Triart C8		12	C8				
Triart Phenyl			Phenyl	1 – 10	20 – 40 °C	50 °C	
Triart PFP			Pentafluorophenyl	No	1 – 8	20 – 40 °C	50 °C
Triart Bio C4		30	C4	Yes	1 – 10	20 – 40 °C	pH 1 – 7: 90 °C pH 7 – 10: 50 °C

### 3. Recommendations for column connections, detector settings, and data processing considerations

- The "PT", "PTH", "PTP" or "WT" at the end of the product code indicates the style of column endfittings (see below for details). Wetted part (Inner surface of column tube) of "PTP" type column is PEEK. Its external tube is made of stainless steel. A special care has to be taken to connect this "PTP" type column to your system. Refer to the "Instructions for column connection YMC-Triart Metal Free Column".

#### Consideration of connector and endfittings



Particle size	The end of the product code	Port depth	Style of endfittings
1.9 $\mu\text{m}$	PT / PTP	ca. 2 mm / 0.09 inch	Parker style
5 $\mu\text{m}$ , 3 $\mu\text{m}$	PTH / PTP	ca. 2 mm / 0.09 inch	Parker style
	WT	ca. 3 mm / 0.13 inch	Waters style

- Tubing must have flat ends and must bottom out in the column endfitting. Tubing must be connected to the column correctly to avoid creating a void between the column frit and tubing, which can cause a leak and result in poor column performance (e.g. peak tailing, loss of theoretical plate number).

- The 1.9  $\mu\text{m}$  column for UHPLC gives high operating pressure compared to conventional 5  $\mu\text{m}$ /3  $\mu\text{m}$  columns. Particular attention should be devoted to the usable maximum pressure of tubing and LC system. Generally, a UHPLC system that has 60 MPa (8,700 psi) or higher pressure tolerance is recommended for use with a 1.9  $\mu\text{m}$  column. We have a connector with a re-adjustable ferrule, which provides flawless sealing at 137 MPa (20,000 psi). Please inquire for related details.
- The extra column volume has a great impact on band spreading. To minimize the influence of band spreading on chromatographic performance, especially when using columns in 2.0 mm I.D. and smaller, the LC system should be optimized as described below.
  - The shortest possible length of tubing with narrow inner diameters (tubing less than 0.15 mm, 0.006 inch I.D. is recommended) should be used for the connection from the injector to the column and from the column to the detector. Make sure not to have a void in the connection.
  - Use a detector equipped with low-volume flow cell designed for the narrow bore column.
  - Use an injector for the narrow bore column and a low-volume sample loop.
- A sampling rate and a detector response (time constant) should be optimized to acquire more than 10 data points across a peak. For UHPLC analysis using 1.9  $\mu\text{m}$  columns, we recommend a sampling rate of about 10 points per second or higher and a detector response of 0.1 s or faster to detect the earliest eluting sharp peak properly.

#### 4. Shipping solvent

100% Acetonitrile. Replace with this solvent for storage. When replacing a mobile phase containing buffer salts/additives, extra care must be taken to prevent salt precipitation.

#### 5. Precautions for use

- The correct direction of the solvent flow is indicated by an arrow on the column identification label.
- Do not disconnect a column from the LC system before the pressure drops to zero.
- The column pressure limit and recommended flow rate are the following.

Particle size	The end of the product code	Pressure limit <sup>*1</sup>	Column I.D. and recommended flow rate <sup>*3</sup> (Mobile phase condition : acetonitrile/aqueous solution)
1.9 $\mu\text{m}$	PT / PTP	100 MPa	2.0/2.1 mm I.D. : 0.2 – 0.8 mL/min 3.0 mm I.D. : 0.4 – 1.6 mL/min
5 $\mu\text{m}$ , 3 $\mu\text{m}$	PTH / PTP	45 MPa <sup>*2</sup>	2.1 mm I.D. : 0.2 mL/min 3.0 mm I.D. : 0.4 mL/min 4.6 mm I.D. : 1.0 mL/min
	WT	Column length of 150 mm and less: 20 MPa Column length of 250 mm: 25 MPa Inner diameter of 10 mm and larger: 10 MPa	2.0 mm I.D. : 0.2 mL/min 3.0 mm I.D. : 0.4 mL/min 4.6 mm I.D. : 1.0 mL/min

\*1 Avoid using a column repeatedly near the pressure limit or abrupt change in pressure to prevent shortening of the column life.

\*2 We recommended using PTH / PTP type columns at a pressure of 30 MPa or less for regular use. Continuous use of the column near the pressure limit may result in short column lifetime.

\*3 Adjust the flow rate appropriately because the pressure changes depending on the column length, temperature, types of organic solvent etc.

- Aqueous or non-aqueous solvents can be used as a mobile phase. Repetitive replacement among solvents with large difference in polarities might degrade the column performance. In general, acetonitrile, methanol and tetrahydrofuran (THF) are recommended for regular use (Please also see the recommendation above under alkaline mobile phase conditions). When using THF as a mobile phase, be mindful of the solvent resistance of your system or tubing (PEEK parts are especially unsuitable for use with THF).
- Recommendations of pH and temperature for column use are shown in the specifications table in section 2. In addition, please note the following.
  - \* Column lifetime varies depending on conditions of use such as pH, temperature and mobile phase composition. In general, usage at higher temperatures, higher concentrations of buffer salts/additives and lower concentrations of organic solvents can shorten the column lifetime.
  - \* When using the column under alkaline conditions for a long time, a column should be used with a low concentration (about 1 – 10 mM) of organic buffer, such as triethylamine or glycine, at a lower temperature (less than 30 °C). The mobile phase containing methanol is expected to give a longer column lifetime under alkaline conditions than other organic solvents will.
- Make sure of the miscibility among the organic solvents and take care to prevent the precipitation of buffer salts/additives to avoid overpressuring the column.

- When possible, the sample should be dissolved in a solvent that is of the same composition as the initial mobile phase. Using a stronger solvent than the initial mobile phase for sample dissolution might result in distorted peak symmetry and degraded resolution.
- To prevent exposure of the column to excessive pressure, the mobile phase and sample solution should be filtered through a 0.2 µm membrane or smaller to remove particulates.
- When using Triart C18 ExRS column, its extremely high hydrophobicity might cause difficulties of replacement or equilibration with mobile phase containing low concentration of organic solvent. The mobile phase should be contained more than 15% of methanol or 10% of other lower polar organic solvent. When replacing methanol aqueous solution with acetonitrile aqueous solution, mobile phase containing less than 20% of acetonitrile may result in irregularities in retention time or peak shapes. In these cases, first replace with 60% acetonitrile aqueous solution, and then replace with the mobile phase.

## 6. Column cleaning (general method)

[After using mobile phase not containing buffer salts/additives]

- Flush the column with solution containing a higher ratio of organic solvent for washing out the compounds that have a great capacity for retention in the column.
- Usable concentration of organic solvent is up to 100%. A cleaning solution containing THF might be effective when removing highly hydrophobic (lipid-soluble) substances that are adsorbed onto the gel.

[After using mobile phase containing buffer salts/additives]

- First replace with a water/organic solution containing no buffer salts or additives (A ratio of water to organic solvent should be set at the same proportions as a mobile phase). Then flush the column in accordance with the method described above.

[General proposals]

- Flushing with 100% water after using the column around the pH limit might shorten column lifetime. Flush the column with water/organic solution as described above.
- Once macromolecules such as proteins or polysaccharides are adsorbed onto the gel, they are hardly removed, even if solvents with high eluting capability are used. To avoid contamination of the column by them, conduct sample pretreatment carefully before introduction into the column. Alternatively, use a guard column.